



PWS_0430035_AC_20190318_Analysis Report
LCRA Environmental Laboratory Services
3505 Montopolis Drive
Austin, TX 78744
Phone: (512)730-6022
Fax: (512)730-6021

March 20, 2019

ROBERT TODD
FROGNOT WSC
PO BOX 400
408 W. FM 545 STE3
Blue Ridge, TX 75424

RE: Final Analytical Report
ELS Sample Q1911123005
FROGNOT WSC

Attn:

Enclosed are the analytical results for the above referenced sample received by LCRA Environmental Laboratory Services. Results reported herein conform to the most current NELAP standards, where applicable, unless otherwise narrated in the body of the report. This final report provides results related only to the sample as received.

Thank you for selecting ELS for your analytical needs. If you have any questions regarding this report, please contact us at (512) 356-6022. We look forward to assisting you again.

Authorized for release by:

Bhanu Acharya

bhanu.acharya@lcra.org

Project Manager

Enclosures





LCRA Environmental Laboratory Services
3505 Montopolis Drive
Austin, TX 78744
(512)730-6022
(512)730-6021

SAMPLE SUMMARY

Workorder: Q1911123

Lab ID	Sample ID	matrix	Date Collected	Date Received
Q1911123005	1937269	Drinking Water	3/18/2019 11:01	3/19/2019 08:15

Report Definitions

RL Reporting Limit
MCL Maximum Contaminant Level
DF Dilution Factor
Qual Qualifiers



LCRA Environmental Laboratory Services
 3505 Montopolis Drive
 Austin, TX 78744
 (512)730-6022
 (512)730-6021

ANALYTICAL RESULTS

Workorder: Q1911123

Lab ID: Q1911123005	Date Received: 3/19/2019 08:15	Matrix: Drinking Water
Sample ID: 1937269	Date Collected: 3/18/2019 11:01	Sample Type: SAMPLE
Facility: EP003	Location: PH	
Sample: TRT-TAP	Client ID: TX0430035	

Parameter	Result	Units	MRL	LOD	MCL	DF	Prepared	By	Analyzed	By	Qual
E300.0, Anions											
Analysis Desc: E300.0, Anions											
Nitrate (as N)	0.0547	mg/L	0.0100	0.00500	10	1			03/19/19 17:57		ML



LCRA Environmental Laboratory Services
3505 Montopolis Drive
Austin, TX 78744
(512)730-6022
(512)730-6021

PARAMETER QUALIFIERS

Workorder: Q1911123

Parameter Qualifiers

J	Analyte Detected Below Quantitation Limit
SH	Spike Recovery High
SL	Spike Recovery Low



LCRA Environmental Laboratory Services
 3505 Montopolis Drive
 Austin, TX 78744
 (512)730-6022
 (512)730-6021

QUALITY CONTROL

QC Batch: 86546
QC Batch Method: E300.0, Anions
Analytical Method: E300.0, Anions

MATRIX SPIKE: 1212895 Original Sample: Q1911097038

Parameter	Units	Original Result	Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	Limit	RPD	Max RPD	Qual
Nitrate (as N)	mg/L	.0842	1	1.08	1.09	100	101	80-120	.922	20	

LABORATORY CONTROL SAMPLE: 1212894

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limit	Qual
Nitrate (as N)	mg/L	1	1.01	101	90-110	

MRL: 1212887

Parameter	Units	Spike Conc.	MRL Result	MRL % Rec	% Rec Limit	Qual
Nitrate (as N)	mg/L	.01	.0115	115	50-150	

METHOD BLANK: 1212893

Parameter	Units	Blank Result	Reporting Limit	Qual
Nitrate (as N)	mg/L	<0.0100	0.0100	

Qualifiers

- SL - Spike Recovery Low
- SH - Spike Recovery High
- R - RPD Outside Recovery Limit